

Quality Assurance Certificate

r-biopharm®



Product: Enzytec™ *Color* IRON (Ferene-S)

REF	Art. No.:	E2300
LOT	Lot:	RBN001
	Expiry:	2021-07

Test	Specifications	Results
Reagent 1 visual check smell pH (25°C ± 1°C)	colourless no characteristic odour 2.0 (1.85 - 2.25)	conforms conforms 2.09
Reagent 2 visual check smell pH (25°C ± 1°C)	yellow-orange no characteristic odour 6.40 (6.20 - 6.60)	conforms conforms 6.46
Reagent 3 visual check smell concentration	clear, colourless no characteristic odour 20 (18 - 22) mg/L	conforms conforms 19.84 – 20.06 mg/L
Control Sample 1	20 (18 - 22) mg/L	19.99 – 20.02 mg/L

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2020-01-21



QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com