

# Quality Assurance Certificate

r-biopharm®



**Product:** Enzytec™ *Color* IRON (Ferene-S)

<b>REF</b>	<b>Art. No.:</b>	E2300
<b>LOT</b>	<b>Lot:</b>	RBP006
	<b>Expiry:</b>	2022-10

Test	Specifications	Results
<b>Reagent 1</b> visual check smell pH (25°C ± 1°C)	colourless no characteristic odour 2.0 (1.85 - 2.25)	conforms conforms 2.07
<b>Reagent 2</b> visual check smell pH (25°C ± 1°C)	yellow-orange no characteristic odour 6.40 (6.20 - 6.60)	conforms conforms 6.41
<b>Reagent 3</b> visual check smell concentration	clear, colourless no characteristic odour 20 (18 - 22) mg/L	conforms conforms 20.02 – 20.07 mg/L
<b>Control Sample 1</b>	20 (18 - 22) mg/L	20.09 – 20.14 mg/L

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

## Quality Control

Date: 2021-04-21

Quality  
Management  
System certified  
• ISO 9001  
• ISO 13485

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com