

Quality Assurance Certificate

r-biopharm®



Product: Enzytec™ Citric Acid

REF	Art. No.:	E1214
LOT	Lot:	200806
	Expiry:	2023-04

Test	Specifications	Results
Part 1: Specific activity	CL \geq 50 U/ml	corresponds
Part 2: pH value	Bottle #1 7.6 - 8.0	corresponds
Part 3: Solubilisation time	Bottle #1 \leq 3 min Bottle #2 \leq 3 min	corresponds corresponds
Performance test	Target values \pm 5%	corresponds

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2021-09-29



QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com