

Quality Assurance Certificate

r-biopharm®



Product: Enzytec™ L-Malic Acid

REF	Art. No.:	E1215
LOT	Lot:	200797
	Expiry:	2023-01

Test	Specifications	Results
Part 1: Specific activity	L-Malate DH \geq 6000 U/ml GOT enzyme \geq 400 U/ml	corresponds corresponds
Part 2: pH value	Bottle #1 9.8 - 10.2	corresponds
Part 3: Solubilisation time	Bottle #2 \leq 3 min	corresponds
Performance test	Target values \pm 5%	corresponds

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2021-06-17

Quality
Management
System certified
• ISO 9001
• ISO 13485

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com