


Quality Assurance Certificate

r-biopharm®



Product: Enzytec™ Acetic Acid

REF	Art. No.:	E1226
LOT	Lot:	200757
	Expiry:	2021-05

Test	Specifications	Results
Part 1: Specific activity	L-Malate DH ≥ 2700 U/ml CS ≥ 700 U/ml Acetyl-CoA synthetase ≥ 22 U/ml	corresponds corresponds corresponds
Part 2: pH value	Bottle #1 8.2 - 8.6	corresponds
Part 3: Solubilisation time	Bottle #2 ≤ 3 min Bottle #4 ≤ 3 min	corresponds corresponds
Performance test	Target values $\pm 5\%$	corresponds

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2019-10-22

Quality
Management
System certified
• ISO 9001
• ISO 13485

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com