Quality Assurance Certificate r-biophorm®





Enzytec™ L-Ascorbic Acid Product:

REF	Art. No.:	E1267
LOT	Lot:	200776
\square	Expiry:	2021-12

Test	Specifications	Results
Part 1: Specific activity	AAO ≥ 600 U/bottle	corresponds
Part 2: pH value	Bottle #1 3.3 - 3.7	corresponds
Part 3: Solubilisation time	Bottle #2 ≤ 3 min	corresponds
Performance test	Target values ±10%	corresponds

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2020-05-19



QAC-Version: 2019-04-01