

# Quality Assurance Certificate

r-biopharm®



**Product:** Enzytec™ Oxalic Acid

|   |                  |         |
|---|------------------|---------|
| <b>REF</b>  | <b>Art. No.:</b> | E2100   |
| <b>LOT</b>  | <b>Lot:</b>      | RBN006  |
|  | <b>Expiry:</b>   | 2021-10 |

| Test   | Specifications   | Results                                      |
|--|--|--|
| <b>Control Sample 1</b>                                      | 0.29 (0.23 – 0.35) mmol/L  | 0.282 – 0.288 mmol/L                         |
| <b>Reagent 1</b><br>visual check<br>smell<br>pH (25°C ± 1°C) | clear, colourless<br>no characteristic odour<br>3.1 ± 0.1                | conforms<br>conforms<br>3.060                |
| <b>Reagent 2</b><br>visual check<br>smell<br>pH (25°C ± 1°C) | clear, colourless<br>no characteristic odour<br>N/A                      | conforms<br>conforms                         |
| <b>Reagent 3</b><br>visual check<br>smell<br>concentration   | clear, colourless<br>no characteristic odour<br>0.5 (0.45 - 0.55) mmol/L | conforms<br>conforms<br>0.499 - 0.500 mmol/L |

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

## Quality Control

Date: 2020-04-20

Quality  
Management  
System certified  
• ISO 9001  
• ISO 13485

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com