

Quality Assurance Certificate

r-biopharm®



Product: Enzytec™ Oxalic Acid

REF	Art. No.:	E2100
LOT	Lot:	RBN023
	Expiry:	2022-04

Test	Specifications	Results
Control Sample 1	0.29 (0.23 – 0.35) mmol/L	0.312 – 0.310 mmol/L
Reagent 1 visual check smell pH (25°C ± 1°C)	clear, colourless no characteristic odour 3.1 ± 0.1	conforms conforms 3.060
Reagent 2 visual check smell pH (25°C ± 1°C)	clear, colourless no characteristic odour N/A	conforms conforms
Reagent 3 visual check smell concentration	clear, colourless no characteristic odour 0.5 (0.45 - 0.55) mmol/L	conforms conforms 0.499 - 0.500 mmol/L

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2020-10-16



QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com