Quality Assurance Certificate r-biophorm®





Enzytec™ Oxalic Acid Product:

REF	Art. No.:	E2100
LOT	Lot:	RBP018
\square	Expiry:	2023-03

Test	Specifications	Results
Control Sample 1	0.24 (0.19 – 0.29) mmol/L	0.272 – 0.272 mmol/L
Reagent 1 visual check smell pH (25°C ± 1°C)	clear, colourless no characteristic odour 3.1 ± 0.1	conforms conforms 3.090
Reagent 2 visual check smell pH (25°C ± 1°C)	clear, colourless no characteristic odour N/A	conforms conforms
Reagent 3 visual check smell concentration	clear, colourless no characteristic odour 0.5 (0.45 - 0.55) mmol/L	conforms conforms 0.502 - 0.500 mmol/L

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2021-09-10



QAC-Version: 2019-04-01