

Quality Assurance Certificate

r-biopharm®



Product: Enzytec™ *Color* COPPER

REF	Art. No.:	E2400
LOT	Lot:	RBP005
	Expiry:	2022-09

Test	Specifications	Results
Reagent 1 visual check smell pH (25°C ± 1°C)	colourless or yellow-orange* no characteristic odour 4.9 (4.2 - 5.1)	conforms conforms 4.90
Reagent 2 visual check smell pH (25°C ± 1°C)	clear, red no characteristic odour 4.9 (4.2 - 5.1)	conforms conforms 4.85
Reagent 3 visual check smell concentration	clear, colourless no characteristic odour 5.0 (4.50 - 5.50) mg/L	conforms conforms 4.998 – 5.003 mg/L
Control Sample 1	5.0 (4.50 – 5.50) mg/L	5.024 – 5.121 mg/L

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2021-03-24



QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com