

Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE Acetic Acid

REF	Art. No.:	RCS4226
LOT	Lot:	2106-3
	Expiry:	2022-07

Content	Lot	Expiry
Reagent 1	2106-3	2022-07
Reagent 2	2106-3	2022-07
Reagent 3	2106-3	2022-07
RFID Card	2106-3	2022-07
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	500.0	452	90.4
Sample 2	125.0	108	86.4
Reagent blank (water sample)	0.0	0	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	125.0	125	100.0
Sample 2	25.0	22	88.0
Reagent blank (water sample)	0.0	0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2021-07-28

Quality Management System certified
• ISO 9001
• ISO 13485

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com