

Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE L-Malic Acid

REF	Art. No.:	RCS4280
LOT	Lot:	2101-2
	Expiry:	2022-03

Content	Lot	Expiry
Reagent 1	2101-2	2022-03
Reagent 2	2101-2	2022-03
RFID Card	2101-2	2022-03
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	1175	1104	94.0
Sample 2	300	306	102.0
Reagent blank (water sample)	0	3	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	235	228	97.0
Sample 2	75	73	97.3
Reagent blank (water sample)	0	0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2021-02-16

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com

Quality Management System certified
• ISO 9001
• ISO 13485