

# Quality Assurance Certificate

r-biopharm®



**Product: RIDA® CUBE Lactose/ D-Galactose**

<b>REF</b>	Art. No.:	RCS4110
<b>LOT</b>	Lot:	2009-1
	Expiry:	2021-12

Content	Lot	Expiry
Reagent 1	2009-1	2021-12
Reagent 2	2009-1	2021-12
RFID Card	2009-1	2021-12
IFU (and other accompanying documents where applicable)		

## Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	2666.7	2464.0	92.4
Sample 2	1000.0	933.0	93.3
Reagent blank (water sample)	0.0	5.0	/

Ultra-Sensitive application (200 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	333.3	305.0	91.5
Sample 2	100.0	93.0	93.0
Reagent blank (water sample)	0.0	0.0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

## Quality Control

Date: 2020-10-15

QAC-Version: 2020-04-27

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com

Quality  
Management  
System certified  
• ISO 9001  
• ISO 13485