


# Quality Assurance Certificate

r-biopharm®



**Product: RIDA® CUBE D-Galactose**

<b>REF</b>	Art. No.:	RCS4120
<b>LOT</b>	Lot:	2009-2
	Expiry:	2021-12

Content	Lot	Expiry
Reagent 1	2009-2	2021-12
Reagent 2	2009-2	2021-12
RFID Card	2009-2	2021-12
IFU (and other accompanying documents where applicable)		

## Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	1666.7	1581.0	94.9
Sample 2	400.0	373.0	93.3
Reagent blank (water sample)	0.0	0.0	/

Ultra-Sensitive application (200 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	153.8	141.0	91.7
Sample 2	40.0	37.0	92.5
Reagent blank (water sample)	0.0	0.0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

## Quality Control

Date: 2020-10-12

QAC-Version: 2020-04-27

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com

Quality  
Management  
System certified  
• ISO 9001  
• ISO 13485