


# Quality Assurance Certificate

r-biopharm®



**Product: RIDA® CUBE D-Glucose**

<b>REF</b>	Art. No.:	RCS4140
<b>LOT</b>	Lot:	1912-1
	Expiry:	2020-12

Content	Lot	Expiry
Reagent 1	1912-1	2020-12
Reagent 2	1912-1	2020-12
RFID Card	1912-1	2020-12
IFU (and other accompanying documents where applicable)		

## Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	1601	1577	98.5
Sample 2	246	240	97.6
Reagent blank (water sample)	0	0	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	393.6	392	99.6
Sample 2	49.2	48	97.6
Reagent blank (water sample)	0	0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

## Quality Control

Date: 2020-01-13

Quality Management System certified  
• ISO 9001  
• ISO 13485

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com