

Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE D-Glucose

REF	Art. No.:	RCS4140
LOT	Lot:	2008-1
	Expiry:	2021-10

Content	Lot	Expiry
Reagent 1	2008-1	2021-10
Reagent 2	2008-1	2021-10
RFID Card	2008-1	2021-10
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	1658.7	1528	92.1
Sample 2	248.5	236	95.0
Reagent blank (water sample)	0	1	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	397.6	389	97.8
Sample 2	49.7	48	96.6
Reagent blank (water sample)	0	0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2020-08-27

Quality Management System certified
• ISO 9001
• ISO 13485

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com