

# Quality Assurance Certificate

r-biopharm®



**Product: RIDA® CUBE D-Glucose/D-Fructose**

<b>REF</b>	Art. No.:	RCS4160
<b>LOT</b>	Lot:	2010-1
	Expiry:	2022-01

Content	Lot	Expiry
Reagent 1	2010-1	2022-01
Reagent 2	2010-1	2022-01
RFID Card	2010-1	2022-01
IFU (and other accompanying documents where applicable)		

## Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	1663.8	1508	90.6
Sample 2	249.5	237	95.0
Reagent blank (water sample)	0	1	Passed

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	399.2	361	90.4
Sample 2	49.9	45	90.2
Reagent blank (water sample)	0	0	Passed

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

## Quality Control

Date: 2020-11-03

Quality Management System certified  
• ISO 9001  
• ISO 13485

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com