

# Quality Assurance Certificate

r-biopharm®



**Product: RIDA® CUBE Sucrose/D-Glucose**

<b>REF</b>	Art. No.:	RCS4180
<b>LOT</b>	Lot:	2000-1
	Expiry:	2021-02

Content	Lot	Expiry
Reagent 1	2001-1	2021-02
Reagent 2	2001-1	2021-02
RFID Card	2001-1	2021-02
IFU (and other accompanying documents where applicable)		

## Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	1448.0	1346.0	93.0
Sample 2	289.6	273.0	94.3
Reagent blank (water sample)	0	0	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	724.0	701.0	96.8
Sample 2	144.8	135.0	93.2
Reagent blank (water sample)	0	0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

## Quality Control

Date: 2020-02-11

Quality Management System certified  
• ISO 9001  
• ISO 13485

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com