

# Quality Assurance Certificate

r-biopharm®



**Product: RIDA® CUBE Sucrose/D-Glucose**

<b>REF</b>	Art. No.:	RCS4180
<b>LOT</b>	Lot:	2007-1
	Expiry:	2021-09

Content	Lot	Expiry
Reagent 1	2007-1	2021-09
Reagent 2	2007-1	2021-09
RFID Card	2007-1	2021-09
IFU (and other accompanying documents where applicable)		

## Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	1396.8	1360	97.4
Sample 2	279.4	271	97.0
Reagent blank (water sample)	0	0	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	698.4	692	99.1
Sample 2	139.7	138	98.8
Reagent blank (water sample)	0	0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

## Quality Control

Date: 2020-07-13

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com

Quality Management System certified  
• ISO 9001  
• ISO 13485