


Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE Sucrose/ D-Glucose/ D-Fructose

REF	Art. No.:	RCS4190
LOT	Lot:	2106-1
	Expiry:	2022-07

Content	Lot	Expiry
Reagent 1	2106-1	2022-07
Reagent 2	2106-1	2022-07
RFID Card	2106-1	2022-07
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	1526.5	1494.0	97.9
Sample 2	305.3	303.0	99.2
Reagent blank (water sample)	0.0	1.0	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	305.3	289.0	94.7
Sample 2	76.3	71.0	93.1
Reagent blank (water sample)	0.0	0.0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2021-06-21



QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com