

Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE L-Lactic acid

REF	Art. No.:	RCS4260
LOT	Lot:	2006-1
	Expiry:	2021-07

Content	Lot	Expiry
Reagent 1	2006-1	2021-07
Reagent 2	2006-1	2021-07
RFID Card	2006-1	2021-07
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	900	861	95.7
Sample 2	140	132	94.3
Reagent blank (water sample)	0	3	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	140	127	90.7
Sample 2	35	32	91.4
Reagent blank (water sample)	0	0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2020-07-07



QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com