


Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE L-Lactic Acid

REF	Art. No.:	RCS4190
LOT	Lot:	2106-2
	Expiry:	2022-07

Content	Lot	Expiry
Reagent 1	2106-2	2022-07
Reagent 2	2106-2	2022-07
RFID Card	2106-2	2022-07
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	1000.2	940.0	94.0
Sample 2	125.0	126.0	100.8
Reagent blank (water sample)	0.0	0.0	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	166.7	155.0	93.0
Sample 2	31.3	29.0	92.7
Reagent blank (water sample)	0.0	1.0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2021-06-18

Quality Management System certified
• ISO 9001
• ISO 13485

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com