


# Quality Assurance Certificate

r-biopharm®



**Product: RIDA® CUBE D/L-Lactic acid**

<b>REF</b>	Art. No.:	RCS4240
<b>LOT</b>	Lot:	1910-4
	Expiry:	2020-11

Content	Lot	Expiry
Reagent 1	1910-4	2020-11
Reagent 2	1910-4	2020-11
RFID Card	1910-4	2020-11
IFU (and other accompanying documents where applicable)		

## Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	900	904	100.4
Sample 2	141	134	95.0
Reagent blank (water sample)	0	0	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	141	144	102.1
Sample 2	35.3	37	104.8
Reagent blank (water sample)	0	0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

## Quality Control

Date: 2019-11-12

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com

Quality Management System certified  
• ISO 9001  
• ISO 13485