

Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE D/L-Lactic acid

REF	Art. No.:	RCS4240
LOT	Lot:	2003-4
	Expiry:	2021-04

Content	Lot	Expiry
Reagent 1	2003-4	2021-04
Reagent 2	2003-4	2021-04
RFID Card	2003-4	2021-04
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	900.0	840	93.3
Sample 2	140.0	133	95.0
Reagent blank (water sample)	0.0	0	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	140.0	131	93.6
Sample 2	35.0	33	94.3
Reagent blank (water sample)	0	0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2020-03-19

Quality Management System certified
• ISO 9001
• ISO 13485

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com