

# Quality Assurance Certificate

r-biopharm®



**Product:** RIDA® CUBE D/L-Lactic acid

<b>REF</b>	Art. No.:	RCS4240
<b>LOT</b>	Lot:	2010-2
	Expiry:	2022-01

Content	Lot	Expiry
Reagent 1	2010-2	2022-01
Reagent 2	2010-2	2022-01
RFID Card	2010-2	2022-01
IFU (and other accompanying documents where applicable)		

## Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	900.0	837.0	93.0
Sample 2	140.0	135.0	96.4
Reagent blank (water sample)	0	2	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	140.0	128.0	91.4
Sample 2	35.0	32.0	91.4
Reagent blank (water sample)	0	0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

## Quality Control

Date: 2020-11-19

Quality Management System certified  
• ISO 9001  
• ISO 13485

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com