

# Quality Assurance Certificate

r-biopharm®



**Product: RIDA® CUBE L-Malic acid**

<b>REF</b>	Art. No.:	RCS4280
<b>LOT</b>	Lot:	2003-3
	Expiry:	2021-04

Content	Lot	Expiry
Reagent 1	2003-3	2021-04
Reagent 2	2003-3	2021-04
RFID Card	2003-3	2021-04
IFU (and other accompanying documents where applicable)		

## Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	1150.0	1128	98.1
Sample 2	147	148	100.7
Reagent blank (water sample)	0	0	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	230.0	227	98.7
Sample 2	36.8	37	100.5
Reagent blank (water sample)	0	0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

## Quality Control

Date: 2020-03-18



QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com