

Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE SO₂-Total

REF	Art. No.:	RCS4600
LOT	Lot:	2102-2
	Expiry:	2022-04

Content	Lot	Expiry
Reagent 1	2102-2	2022-04
Reagent 2	2102-2	2022-04
RFID Card	2102-2	2022-04
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	250.0	237.9	95.2
Sample 2	50.0	49.0	98.0
Reagent blank (water sample)	0.0	1.0	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	50.0	49.4	98.8
Sample 2	15.0	13.6	90.7
Reagent blank (water sample)	0.0	1.0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2021-03-09

Quality Management System certified
• ISO 9001
• ISO 13485

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com