

# Quality Assurance Certificate

r-biopharm®



**Product: RIDA® CUBE SO<sub>2</sub>-Total**

<b>REF</b>	Art. No.:	RCS4600
<b>LOT</b>	Lot:	2109-1
	Expiry:	2022-10

Content	Lot	Expiry
Reagent 1	2109-1	2022-10
Reagent 2	2109-1	2022-10
RFID Card	2109-1	2022-10
IFU (and other accompanying documents where applicable)		

## Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	250.0	249.7	99.9
Sample 2	50.0	51.3	102.6
Reagent blank (water sample)	0.0	1.0	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	50.0	49.6	99.2
Sample 2	15.0	14.6	97.3
Reagent blank (water sample)	0.0	0.0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

## Quality Control

Date: 2021-09-30



QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com