

# Quality Assurance Certificate

r-biopharm®



**Product: RIDA® CUBE Lactose/D-Galactose**

<b>REF</b>	Art. No.:	RCS4110
<b>LOT</b>	Lot:	2104-2
	Expiry:	2022-05

Content	Lot	Expiry
Reagent 1	2104-2	2022-05
Reagent 2	2104-2	2022-05
RFID Card	2104-2	2022-05
IFU (and other accompanying documents where applicable)		

## Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	2899.8	2622.0	90.4
Sample 2	1000.0	951.0	95.1
Reagent blank (water sample)	0.0	1.0	/

Ultra-Sensitive application (200 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	362.5	331.0	91.3
Sample 2	100.0	91.0	91.0
Reagent blank (water sample)	0.0	0.0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

## Quality Control

Date: 2021-04-13

Quality  
Management  
System certified  
• ISO 9001  
• ISO 13485

QAC-Version: 2020-04-27

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com