

Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE D-Galactose

REF	Art. No.:	RCS4120
LOT	Lot:	2104-3
	Expiry:	2022-05

Content	Lot	Expiry
Reagent 1	2104-3	2022-05
Reagent 2	2104-3	2022-05
RFID Card	2104-3	2022-05
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	1666.7	1534.0	92.0
Sample 2	500.0	478.0	95.6
Reagent blank (water sample)	0.0	1.0	/

Ultra-Sensitive application (200 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	166.7	154.0	92.4
Sample 2	20.0	19.0	95.0
Reagent blank (water sample)	0.0	0.0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2021-04-12

Quality
Management
System certified
• ISO 9001
• ISO 13485

QAC-Version: 2020-04-27

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com