

Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE D-Galactose

REF	Art. No.:	RCS4120
LOT	Lot:	1907-2
	Expiry:	2020-10

Content	Lot	Expiry
Reagent 1	1907-2	2020-10
Reagent 2	1907-2	2020-10
RFID Card	1907-2	2020-10
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	1666.7	1723.5	103.4
Sample 2	400.0	407.4	101.9
Reagent blank (water sample)	0.0	0.0	/

Sensitive application (200 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	153.8	155.6	101.2
Sample 2	40.0	39.0	97.5
Reagent blank (water sample)	0.0	0.0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2019-07-30



QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com