

# Quality Assurance Certificate

r-biopharm®



**Product:** RIDA® CUBE D/L-Lactic Acid

<b>REF</b>	Art. No.:	RCS4240
<b>LOT</b>	Lot:	2104-1
	Expiry:	2022-04

Content	Lot	Expiry
Reagent 1	2104-1	2022-04
Reagent 2	2104-1	2022-04
RFID Card	2104-1	2022-04
IFU (and other accompanying documents where applicable)		

## Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	900.0	813.0	90.3
Sample 2	141.0	138.0	97.9
Reagent blank (water sample)	0.0	2.0	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	141.0	132.0	93.6
Sample 2	35.3	32.0	90.7
Reagent blank (water sample)	0.0	0.0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

## Quality Control

Date: 2021-04-12

Quality Management System certified  
• ISO 9001  
• ISO 13485

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com