

Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE SO₂-Free

REF	Art. No.:	RCS4610
LOT	Lot:	2007-2
	Expiry:	2021-09

Content	Lot	Expiry
Reagent 1	2007-2	2021-09
Reagent 2	2007-2	2021-09
RFID Card	2007-2	2021-09
IFU (and other accompanying documents where applicable)		

Release test

Ultra-Sensitive application (200 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	24.0	23.0	95.8
Sample 2	8.0	8.0	100.0
Reagent blank (water sample)	0	0	/

„Creep reaction“ (200 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	24.0	25.0	104.2
Sample 2	8.0	8.0	100.0
Reagent blank (water sample)	0	0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2020-07-20

QAC-Version: 2020-09-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com

Quality Management System certified
• ISO 9001
• ISO 13485