

Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE SO₂-Free

REF	Art. No.:	RCS4610
LOT	Lot:	2102-4
	Expiry:	2022-05

Content	Lot	Expiry
Reagent 1	2102-4	2022-05
Reagent 2	2102-4	2022-05
RFID Card	2102-4	2022-05
IFU (and other accompanying documents where applicable)		

Release test

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	24.0	21.8	90.8
Sample 2	12.0	11.0	91.7
Reagent blank (water sample)	0	0.4	passed

„Creep reaction“ (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	24.0	21.8	90.8
Sample 2	12.0	11.5	95.8
Reagent blank (water sample)	0	0.8	passed

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2021-04-01

Quality Management System certified
• ISO 9001
• ISO 13485

QAC-Version: 2021-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com