

Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE Sucrose/D-Glucose

REF	Art. No.:	RCS4180
LOT	Lot:	2107-1
	Expiry:	2022-08

Content	Lot	Expiry
Reagent 1	2107-1	2022-08
Reagent 2	2107-1	2022-08
RFID Card	2107-1	2022-08
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	2900	2793	96.3
Sample 2	290	266	91.7
Reagent blank (water sample)	0	0	passed

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	725	708	97.7
Sample 2	145	133	91.7
Reagent blank (water sample)	0	0	passed

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2021-08-19

Quality Management System certified
• ISO 9001
• ISO 13485

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com