

Quantitative determination of high molecular weight β -glucan from barley in malt-mash and wort

Art. No. E3500 test combination for 40 determinations

Art. No. E3550 test combination for 160 determinations

For *in vitro* use only
Store between 2 - 8 °C

Approval: EBC 4.16.3, MEBAK 3.1.4.9.2, ASBC

1. Test principle

Colorimetric assay for the quantitative determination of high molecular weight β -glucan from barley in malt-mash and wort.

The test has been recognized under EBC 4.16.3 (high molecular weight β -glucan content of malt wort: spectrophotometric method), MEBAK 3.1.4.9.2 (high molecular β -Glucan colorimetric method) and ASBC (β -glucan in wort – spectrophotometric method).

2. Reagents

2.1. Content & composition

The test is suitable for manual processing. With manual processing, the reagents are sufficient for 40 or 160 determinations.

E3500 test combination for 40 determination:

- Solution A: 1 x 125 mL
- Calibrator set: 6 x 1.0 mL (0, 100, 200, 300, 400, 500 mg β -glucan/L)

E3550 test combination for 160 determination:

- Solution A: 4 x 125 mL
- Calibrator set: 6 x 1.0 mL (0, 100, 200, 300, 400, 500 mg β -glucan/L)

2.2. Reagent preparation

The reagents are ready-to-use and be allowed to reach room temperature (20 - 25 °C) before use. Do not interchange components between kits of different batches.

2.3. Storage & stability

The reagents are stable until the end of the month of the indicated shelf life (see label) even after opening at 2 - 8 °C if handled properly. Do not freeze reagents.

2.4. Safety & disposal

The general safety rules for working in chemical laboratories should be applied. Do not swallow! Avoid contact with skin and mucous membranes.

This kit may contain hazardous substances. For hazard notes on the contained substances, please refer to the appropriate safety data sheets (SDS) for this product. After use, the reagents can be disposed of with the laboratory waste. Packaging materials may be recycled.

3. Sample preparation

- The samples should be brought to room temperature before measurement.
- The method for sample preparation has been described under EBC 4.5.1 (extract of malt: congress mash). It is mandatory to follow it if the present β -glucan kit should be used according to EBC 4.16.3.
- For the determination of the whole content of β -glucans (soluble and insoluble parts), it is recommended to heat the samples for 15 min at 80 °C to solubilize gel-formed fractions of β -glucans; after cooling down to room temperature, a centrifugation or filtration step might be necessary before testing the sample.

4. Assays performance

Wavelength: 550 nm (\pm 5 nm)
Light path: 1.00 cm (glass; plastic)
Temperature: 18 - 30 °C (temperature dependent calibration necessary)
Method: end-point measurement
Incubation time: 30 minutes
Measurement: against air (without cuvette) or water
Measuring range: 100 - 500 mg/L β -glucan

Pipette into cuvette*:	Calibrator (1 cuvette each)	Sample
Calibrator set (1 - 6)	0.200 mL	-
Sample	-	0.200 mL
Solution A	3.000 mL	3.000 mL
Mix, incubate for 30 min at room temperature (20 - 25 °C). Measure absorption (A_1) at 550 nm.		

* Before pipetting a sample or calibrator into the cuvette, first rinse the pipette tip once or twice with the calibrator / sample.

5. Calculation of results

Calculation of sample solutions – total concentration of β -glucan

Plot the calibration curve using the absorbance of the calibrators 1 to 6. The β -glucan content of the samples can be directly read from the curve or calculated from the resulting equation (linear regression). An excel sheet is available on request.

Example with typical absorbance values:

Calibrator	β -glucan (mg/L)	A_1	minus blank
1	0	0.850	0.000
2	100	0.973	0.123
3	200	1.081	0.231
4	300	1.184	0.334
5	400	1.280	0.430
6	500	1.377	0.527

The specific values of each lot can be seen in the corresponding certificate of analysis.

6. Performance data

Linearity, measuring range & sensitivity

Linearity is given up to 500 mg/L β -glucan. The limit of detection (LoD) and limit of quantification (LoQ) were determined according to method DIN 32645:2008-11, using buffered aqueous solutions. This results in an LoD of 7 mg/L and an LoQ of 19 mg/L.

7. Supporting documents

On request, we offer the following document:

- Enzytec™ Liquid Excel template for result calculation

Safety data sheets (SDS) und certificates of analysis (CoA) are available in digital form under the following link

<https://eifu.r-biopharm.com/>



8. Services & technical support

On request, we offer the following services:

- Customized troubleshooting
- Data & results analysis
- Customer workshops & webinars
- Automation: application support and technical service

9. Disclaimer

This information corresponds to our present state of technology and provides information on our products and their uses. R-Biopharm makes no warranty of any kind, either expressed or implied, except that the materials from which its products are made are of standard quality. Defective products will be replaced. There is no warranty of merchantability of this product, or of the fitness of the product for any purpose. R-Biopharm shall not be liable for any damages, including special or consequential damage, or expense arising directly or indirectly from the use of this product.