

Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE Acetic Acid

REF	Art. No.:	RCS4226
LOT	Lot:	2111-1
	Expiry:	2022-12

Content	Lot	Expiry
Reagent 1	2111-1	2022-12
Reagent 2	2111-1	2022-12
Reagent 3	2111-1	2022-12
RFID Card	2111-1	2022-12
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	500	478	95.6
Sample 2	125	107	85.6
Reagent blank (water sample)	0	2	passed

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	125	123	98.4
Sample 2	25	22	88.0
Reagent blank (water sample)	0	0	passed

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2021-12-10

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com

Quality Management System certified
• ISO 9001
• ISO 13485