

Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE D-Glucose/D-Fructose

REF	Art. No.:	RCS4160
LOT	Lot:	2109-3
	Expiry:	2022-11

Content	Lot	Expiry
Reagent 1	2109-3	2022-11
Reagent 2	2109-3	2022-11
RFID Card	2109-3	2022-11
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	1666.7	1660	99.6
Sample 2	250.0	255	102.0
Reagent blank (water sample)	0.0	0	passed

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	400.0	403	108.8
Sample 2	50.0	50	100.0
Reagent blank (water sample)	0.0	0	passed

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2021-10-05



QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com