Quality Assurance Certificate r-biophorm®





RIDA®CUBE Ammonia **Product:**

REF A	rt. No.:	RCS4390
LOT Lo	ot:	2010-3
	piry:	2022-01

Content	Lot	Expiry			
Reagent 1	2010-3	2022-01			
Reagent 2	2010-3	2022-01			
RFID Card	2010-3	2022-01			
IFU (and other accompanying documents where applicable)					

Release test

Basic application (20 μl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	100	94	93.8
Sample 2	20	19	94.0
Reagent blank (water sample)	0	0	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	25	23	90.4
Sample 2	5	5	94.0
Reagent blank (water sample)	0	0	1

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2020-11-20



QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.