

# Quality Assurance Certificate

r-biopharm®



**Product: RIDA® CUBE Amino Nitrogen**

<b>REF</b>	Art. No.:	RCS4630
<b>LOT</b>	Lot:	2109-2
	Expiry:	2022-12

Content	Lot	Expiry
Reagent 1	2109-2	2022-12
Reagent 2	2109-2	2022-12
RFID Card	2109-2	2022-12
IFU (and other accompanying documents where applicable)		

## Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	125	121.9	97.5
Sample 2	50	46.9	93.8
Reagent blank (water sample)	0	0	passed

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	25	24.3	97.2
Sample 2	10	10.1	101
Reagent blank (water sample)	0	0	passed

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

## Quality Control

Date: 2021-09-14

Quality Management System certified  
• ISO 9001  
• ISO 13485

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com