

Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE D-Glucose

REF	Art. No.:	RCS4140
LOT	Lot:	1804-2
	Expiry:	2019-04

Content	Lot	Expiry
Reagent 1	1804-2	2019-04
Reagent 2	1804-2	2019-04
RFID Card	1804-2	2019-04
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	1992.4	1936	97.2
Sample 2	248.0	264	106.5
Reagent blank (water sample)	0	9	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	396.8	392	98.8
Sample 2	49.6	49	98.8
Reagent blank (water sample)	0	0	/

R-Biopharm AG, Darmstadt, Germany, certifies that this batch has been approved by the Quality Control and conforms to specifications.

R-Biopharm AG
Quality Control

Date: 2018-05-14

Quality Management System certified
• ISO 9001
• ISO 13485

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com

2017-01-01