

Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE SO₂-Total

REF	Art. No.:	RCS4600
LOT	Lot:	1901-3
	Expiry:	2020-02

Content	Lot	Expiry
Reagent 1	1901-3	2020-02
Reagent 2	1901-3	2020-02
RFID Card	1901-3	2020-03
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	250.0	242.7	97.1
Sample 2	50.0	49.7	99.4
Reagent blank (water sample)	0	3	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	50.0	47.0	94.0
Sample 2	20.0	18.5	92.5
Reagent blank (water sample)	0	0	/

R-Biopharm AG, Darmstadt, Germany, certifies that this batch has been approved by the Quality Control and conforms to specifications.

R-Biopharm AG
Quality Control

Date: 2018-02-13

Quality Management System certified
• ISO 9001
• ISO 13485

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com