

Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE SO₂-Total

REF	Art. No.:	RCS4600
LOT	Lot:	1910-2
	Expiry:	2020-10

Content	Lot	Expiry
Reagent 1	1910-2	2020-10
Reagent 2	1910-2	2020-10
RFID Card	1910-2	2020-10
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	250.0	249.3	99.7
Sample 2	50.0	48.0	96.0
Reagent blank (water sample)	0.0	0.0	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	50.0	49.0	98.0
Sample 2	20.0	18.9	94.5
Reagent blank (water sample)	0.0	0.0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2019-10-21



QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com