

Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE D-Glucose

REF	Art. No.:	RCS4140
LOT	Lot:	1711-4
	Expiry:	2019-03

Content	Lot	Expiry
Reagent 1	1711-4	2019-03
Reagent 2	1711-4	2019-03
RFID Card	1711-4	2019-03
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	1992.4	1948.0	97.8
Sample 2	246.5	262.0	106.3

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	394.4	393.0	99.6
Sample 2	49.3	54.0	109.5

R-Biopharm AG, Darmstadt, Germany, certifies that this batch has been approved by the Quality Control and conforms to specifications.

R-Biopharm AG
Quality Control

Date: 2017-12-04

Quality
Management
System certified
• ISO 9001
• ISO 13485

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com

2017-01-01