



INSTITUT FÜR ANGEWANDTE LABORANALYSEN GMBH

QuickGEN First-Beer PCR Kit

P1 Taqman[®] Screening
(without yeast detection)

Real-time PCR Kit zum Nachweis von
bierschädlichen Bakterien

real-time PCR Kit for detection of
beer spoilage bacteria

Art. Nr. / Order No.: QPP1ToH 0048 low

Version 02/18

GEN-IAL GmbH
Tel: 0049 2241 2522980
Fax:0049 2241 2522989
info@gen-ial.de
www.gen-ial.de

QuickGEN First-Beer PCR Kit

P1 TaqMan[®] Screening ohne Hefen

Art. Nr.: QPP1ToH 0048 low

(Version 02/18)

1. Beispiele erfolgreich nachgewiesener Mikroorganismen

Bakterien

- Lactobacillus spp.
- L. acetotolerans
- L. acidophilus
- L. amylolyticus
- L. amylovorus
- L. backii
- L. brevis
- L. brevisimilis
- L. buchneri
- L. casei / paracasei
- L. collinoides / paracollinoides
- L. coryniformis ssp. coryniformis
- L. coryniformis ssp. torquens
- L. curvatus
- L. delbrückii ssp. delbrückii/ lactis
- L. harbinensis
- L. helveticus
- L. johnsonii
- L. lindneri
- L. malefermentans
- L. parabrevis
- L. parabuchneri (frigidus)
- L. paucivorans
- L. pentosus
- L. perolens
- L. plantarum
- L. paraplantarum
- L. rossiae
- Lactococcus lactis
- Megasphaera spp.
- Micrococcus kristinae
- Pectinatus spp.
- Pectinatus cerevisiophilus
- Pectinatus frisingensis
- Pectinatus haikarae
- Pectinatus ssp.
- Pediococcus spp.
- Pediococcus acidilactici
- Pediococcus damnosus
- Pediococcus dextrinicus
- Pediococcus inoptinatus
- Pediococcus pentosaceus
- Pediococcus parvulus
- Pediococcus claussenii

2. Verwendungszweck

Nachweis von bakteriellen Kontaminationen in Bier und Biermischgetränken.

3. Testprinzip

Die Detektion erfolgt mittels Fluoreszenzmessung durch das Hydrolysesondenformat (TaqMan®). Durch hot-start-PCR plus doppelt markierten sequenzspezifischen Sonden (FAM/DQ; ROX/DQ) wird bei korrekter Hybridisierung an die Zielsequenz in der Extensions-Phase ein messbares Fluoreszenzsignal definierter Wellenlänge emittiert. Eine Inhibitionskontrolle (HEX/DQ) wird zusammen mit der spezifischen Sequenz in einem Reaktionsgefäß amplifiziert, um falsch negative Ergebnisse durch Inhibition auszuschließen. Die PCR-Systeme enthalten dUTP, welches bei der Elongation zum Teil das dTTP ersetzt. Die Verwendung von Uracil-N-Glycosylase (UNG) eliminiert alle dUMP enthaltenden Amplikons, die aus eventuellen Kontaminationen früherer PCRs stammen könnten. (Das UNG Enzym ist in diesem Kit nicht enthalten).

4. Packungsinhalt

Mit den Reagenzien können 48 Bestimmungen durchgeführt werden:

1 x Premix QPP1ToH	weißer Deckel
48 x QPP1ToH mix (lyophilisiert, inkl. IK-DNA)	tube strips
1 x ddH ₂ O	farbloser Deckel
1 x Control-DNA (lyophilisiert)	gelber Deckel

5. Lagerung

Die Control-DNA wird lyophilisiert geliefert und muss vor Gebrauch in ddH₂O gelöst werden (siehe Punkt 7.1).

Den lyophilisierten QPP1ToH mix und die lyophilisierte Control-DNA nicht einfrieren.

Alle Reagenzien bis auf den Premix bei 4 – 8 °C lagern.

Den Premix bei -20 °C lagern. Wiederholtes Auftauen und Einfrieren (> 3x) sollte vermieden werden, da dadurch die Sensitivität verringert wird. Bei unregelmäßigem Gebrauch sollte deshalb der Premix aliquotiert werden.

Der QPP1ToH mix enthält die fluoreszenzmarkierten Sonden und ist lichtempfindlich. Aus diesem Grund sollte er nicht unnötigem Lichteinfall ausgesetzt werden.

Alle Reagenzien sind bei korrekter Lagerung 6 Monate haltbar.

6. Zusätzlich erforderliches Material

6.1. Geräte

Real-time PCR Gerät für low profile PCR tubes clear
Zentrifuge passend für 1,5 – 2,0 mL Reaktionsgefäße
Zentrifuge für Strips
Pipetten
„Vortex“

6.2. Reagenzien und Verbrauchsmaterialien

steriles, doppelt-destilliertes oder deionisiertes Wasser (ddH₂O)
passende, sterile Filterspitzen (Filtertips)
optional: Uracil N-Glycosylase (0,01 U/μL PCR-Reaktion)

7. PCR

7.1. PCR-Ansatz

Vor der ersten Benutzung muss die lyophilisierte Control-DNA kurz zentrifugiert und in ddH₂O gelöst werden:

- die lyophilisierte Control-DNA in 55 µL ddH₂O aufnehmen
- 15 Minuten lösen

Alle PCR Komponenten vor Gebrauch gut mischen und kurz abzentrifugieren.

PCR-Ansatz pro Probe:

PCR-Komponenten	Menge (µL)
Premix QPP1T	15,0
Proben-DNA	5,0*
Gesamtvolumen	20,0


* bei Verwendung des Simplex Easy DNA Kits: 2,5 µL DNA einsetzen und 2,5 µL steriles ddH₂O hinzufügen

1. Je 15,0 µL Premix in die einzelnen PCR-Reaktionsgefäße füllen
2. 5 µL Proben-DNA zu den vorbereiteten PCR Gefäßen geben, für die PCR-Positivkontrolle 5 µL Control-DNA, für die Extraktionskontrolle 5 µL und für die PCR-Negativkontrolle* 5 µL steriles ddH₂O pipettieren (Pipettenspitzen unbedingt nach jeder Probe wechseln).
3. Die PCR-Reaktionsgefäße sofort verschließen und kurz zentrifugieren.
4. Die PCR-Gefäße ins PCR-Gerät stellen und den Lauf starten.

Sehr wichtig: * Die PCR-Negativkontrolle bitte auf jeden Fall mit 5 µL ddH₂O auffüllen, um unspezifische Amplifikationen zu verhindern.

Zügig arbeiten, Lichteinfall und Erwärmung der Ansätze vermeiden.

Analysediagramm



Diese Angaben entsprechen dem heutigen Stand unserer Kenntnisse und sollen über unsere Produkte und deren Anwendungsmöglichkeiten informieren. Sie haben nicht die Bedeutung, bestimmte Eigenschaften der Produkte oder deren Eignung für einen konkreten Einsatzzweck zuzusichern. GEN-IAL übernimmt keine Gewährleistung, außer für die standardisierte Qualität der Reagenzien. Defekte Produkte werden ersetzt. Darüber hinaus gehende Ansprüche für direkte oder indirekte Schäden oder Kosten aus der Nutzung der Produkte entstehen nicht.

Rechtlicher Hinweis: Die Polymerase-Kettenreaktion (PCR) ist patentrechtlich geschützt und somit lizenzpflichtig. Sie ist im Besitz der Hoffman-La Roche Inc. Diese Produktinformation versteht sich nicht als Autorisierung oder Lizenzierung, die PCR-Methode kommerziell anzuwenden.

QuickGEN First-Beer PCR Kit

P1 TaqMan[®] Screening without yeast

Order No.: QPP1ToH 0048 low

(Version 02/18)

1. Examples of successful detected microorganisms

Bacteria

- Lactobacillus spp.
 - L. acetotolerans
 - L. acidophilus
 - L. amylolyticus
 - L. amylovorus
 - L. backii
 - L. brevis
 - L. brevisimilis
 - L. buchneri
 - L. casei / paracasei
 - L. collinoides / paracasei
 - L. coryniformis ssp. coryniformis
 - L. coryniformis ssp. torquens
 - L. curvatus
 - L. delbrückii ssp. delbrückii/ lactis
 - L. harbinensis
 - L. helveticus
 - L. johnsonii
 - L. lindneri
 - L. malefermentans
 - L. parabrevis
 - L. parabuchneri (frigidus)
 - L. paucivorans
 - L. pentosus
 - L. perolens
 - L. plantarum
 - L. paraplantarum
 - L. rossiae
- Lactococcus lactis
- Megasphaera spp.
- Micrococcus kristinae
- Pectinatus spp.
 - Pectinatus cerevisiiphilus
 - Pectinatus frisingensis
 - Pectinatus haikarae
 - Pectinatus ssp.
- Pediococcus spp.
 - Pediococcus acidilactici
 - Pediococcus damnosus
 - Pediococcus dextrinicus
 - Pediococcus inoptinatus
 - Pediococcus pentosaceus
 - Pediococcus parvulus
 - Pediococcus claussenii

2. Intended use

Qualitative detection of bacterial contaminations in beer and beer mixes.

3. Test principle

The TaqMan[®] real-time PCR is based on hot-start-PCR and sequence-specific dual labelled probes (FAM/DQ; ROX/DQ) which, when accurately hybridised, emit a measurable fluorescent signal of a defined wavelength in the extension phase. The increase of signal is continuously measured in a real-time PCR detection instrument. To avoid false negative PCR-results an Inhibition Control (HEX/DQ) is amplified together in one reaction vessel with the specific sequence. The system contains dUTP. Optional: the use of Uracil-N-Glycosylase will eliminate any contamination with Uracil containing amplicons from former PCRs (the enzyme is not part of this kit).

4. Kit contents

The kit contains sufficient reagents for 48 PCR reactions:

1 x Premix QPP1ToH	white cap
48 x QPP1ToH mix (freeze-dried, incl. IC-DNA)	tube strips
1 x ddH ₂ O	colourless cap
1 x Control-DNA (freeze-dried)	yellow cap

5. Storage conditions

The Control-DNA is freeze-dried and has to be solved in ddH₂O prior to use (see 6.1).

Do **not** freeze the lyophilized QPP1ToH mix and lyophilized Control-DNA.

All PCR reagents, except the Premix should be stored at 4 – 8 °C (39 – 46 °F).

Keep Premix for storage at - 20 °C (- 4 °F). Avoid loss of sensitivity by repeating freezing and thawing more than 3 times. For irregular use aliquot the Premix.

The QPP1ToH mix contains the fluorescent labelled probes and should be handled light protected.

All reagents are stable for 6 months, if they are stored correctly

6. Materials required, but not provided

6.1. Instruments

Real-time PCR instrument for low profile tubes clear

Centrifuge for 1.5 – 2.0 mL reaction tubes

Centrifuge for strips

Pipettes

“Vortex”

6.2. Reagents and plastic ware

sterile ddH₂O

sterile filter tips

optional: Uracil N-Glycosylase (0.01 U/μL added to the PCR reaction)

7. PCR

7.1. PCR-Setup

When using the kit for the first time, the freeze-dried Control-DNA has to be shortly centrifuged and carefully resolved:

- add 55 μL sterile ddH₂O to the freeze-dried Control-DNA
- after 15 minutes mix well

Before every use thoroughly mix all PCR components and centrifuge briefly.

PCR-reaction Setup:

PCR-Components	volume (μL)
Premix QPP1T	15.0
Sample-DNA	5.0*
Total volume	20.0

* if using the Simplex easy DNA kit: add 2.5 μL DNA and 2.5 μL sterile ddH₂O

1. Pipette 15 μL of the Premix into each PCR-tube, making sure that, prior to the first filling, the tip has been moistened.
2. Add 5 μL sample DNA, add 5 μL of the Control-DNA for the positive control reaction, add 5 μL of the extraction control and 5 μL of ddH₂O for the negative control* reaction. Use a fresh tip with each DNA filling.
3. Close the tubes immediately and centrifuge them shortly.
4. Place the tubes in the PCR-machine and start run.


Very important: * Please fill up the negative control with 5 μL ddH₂O to avoid unspecific amplification.

Work swiftly to avoid warming up and keep away from light.

be positive.

A sample is negative, if there is no detectable fluorescence increase in the different channels and the positive controls have a positive fluorescence signal. The negative controls show no amplification. The Inhibition Control in HEX- or JOE-channel has to be positive in the sample and in the negative controls, a false negative result due to inhibitory effects is then excluded.

analysis flowchart



Note:

The polymerase-chain reaction (PCR) is protected by patents and requires a licence from Hoffmann-LaRoche Inc.. The provided product does not authorise the purchaser for the commercial use of this method. GEN-IAL makes no warranty of any kind, either expressed or implied, except that the materials from which its products are made are of standard quality. If any materials are defective, GEN-IAL will provide a replacement product. There is no warranty of merchantability of this product, or of the fitness of the product for any purpose. GEN-IAL shall not be liable for any damages, including special or consequential damage, or expense arising directly or indirectly from the use of this product.