

Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE Lactose/ D-Galactose

REF	Art. No.:	RCS4110
LOT	Lot:	1907-1
	Expiry:	2020-10

Content	Lot	Expiry
Reagent 1	1907-1	2020-10
Reagent 2	1907-1	2020-10
RFID Card	1907-1	2020-10
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	2666.7	2684.5	100.7
Sample 2	750.0	783.7	104.5
Reagent blank (water sample)	0.0	0.0	/

Sensitive application (200 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	375.0	354.1	94.4
Sample 2	75.0	77.9	103.9
Reagent blank (water sample)	0.0	0.0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2019-08-01

Quality Management System certified
• ISO 9001
• ISO 13485

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com