

# Quality Assurance Certificate

r-biopharm®



**Product: RIDA® CUBE Lactose/D-Glucose**

<b>REF</b>	Art. No.:	RCS4130
<b>LOT</b>	Lot:	2001-2
	Expiry:	2021-02

Content	Lot	Expiry
Reagent 1	2001-2	2021-02
Reagent 2	2001-2	2021-02
RFID Card	2001-2	2021-02
IFU (and other accompanying documents where applicable)		

## Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	2666.7	2483	93.1
Sample 2	1643	1536	93.5
Reagent blank (water sample)	0	10	/

Ultra-Sensitive application (200 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	821.5	793	96.5
Sample 2	164.3	158	96.2
Reagent blank (water sample)	0	0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

## Quality Control

Date: 2020-03-17



QAC-Version: 2020-04-27

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com